



MEASUREMENT SYSTEM FOR HIGH CURRENT SHUNTS DC CHARACTERIZATION AT CMI

Nováková Zachovalová, V., Šíra, M., Streit, J. and Indra, L.

Czech Metrology Institute, Okružní 31, 638 00 Brno, Czech Republic, vnovakovazachovalova@cmi.cz

Abstract

This paper describes a measurement system for DC characterization (direct resistance, temperature coefficient measurements and power coefficient measurements) of low resistance standards (below 0.1 Ω) in wide current level (from 5 A up to 100 A) with uncertainties at ppm level.

Measurement set up

Method: measurement of ratio of output voltages of the tested and the reference standard by dual channel multimeter. Thus, the resistance of tested standard is

calculated:
$$R_x = \frac{U_x}{U_E} \cdot R_E$$

Equipment:

DMM FLUKE 8508A, transconductance amplifier C&H 8100.

Set of working standards

Verification: by comparison with primary resistance standards calibrated in QHS laboratory

Parameters: oil filled and placed in the oil bath (with uncertainty contribution two hundredths of degrees of Celsius)

Nominal value	Current level	Type	Producer
0.02 Ω	5A - 20A	1682	Tinsley
0.01 Ω	10A - 20A	RN I	Metra
0.001 Ω	30A - 100A	RN I	Metra

Power and temperature coefficient measurement

The power coefficient is easily to calculate as:

PCR measured and calculated **in current range of 50% - 100% of nominal current**

The temperature coefficient is easily to calculate as:

TCR measured **in temperature range from 18 °C up to 30 °C at 1/10 of nominal current**



Equipment for TCR measurement:

air thermostatic box
(temp. contribution 0.15°C, humidity 50 %)

standards:
0.02 Ω ; 0.1 Ω (Tinsley oil filled); PT100

Uncertainties calculations

DCR:

Combined uncertainty:
$$u^2(\delta R_x) = c_R^2 u^2(\delta R) + c_E^2 u^2(\delta R_E) + c_{r_c}^2 u^2(\delta r_c) + c_{Edrift}^2 u^2(\delta R_{Edrift}) + c_{Etemp}^2 u^2(\delta R_{Etemp}) + c_{Xtemp}^2 u^2(\delta R_{Xtemp})$$

where

δ_{r_c} is influence of non complete elimination of thermoelectric voltages and multimeter stability,

δ_{Edrift} is influence of resistance value change since last calibration,

δ_{Etemp} and δ_{Xtemp} is contribution of temperature,

c_R , c_E , c_{Rc} , c_{Edrift} , c_{Etemp} , c_{Xtemp} are coefficients of sensitivity.

TCR:

Combined uncertainty:
$$u^2(\delta TC_R) = c_{\Delta R}^2 u^2(\delta \Delta R) + c_{\Delta T}^2 u^2(\delta \Delta T)$$

with $c_{\Delta R} = TC_R / \Delta R$, $c_{\Delta T} = TC_R / \Delta T$,

$$u^2(\delta \Delta T) = u^2(\delta T_2) + u^2(\delta T_1), u^2(\delta \Delta R) = u^2(\delta R_1) + u^2(\delta R_2).$$

PCR:

Combined uncertainty:
$$u^2(\delta PC_R) = c_{\Delta R}^2 u^2(\delta \Delta R) + c_{\Delta W}^2 u^2(\delta \Delta W)$$

with $c_{\Delta R} = PC_R / \Delta R$, $c_{\Delta W} = PC_R / \Delta P$,

$$u^2(\delta \Delta R) = u^2(\delta R_1) + u^2(\delta R_2), u^2(\delta \Delta P) = u^2(\delta P_1) + u^2(\delta P_2),$$

$$u^2(\delta P_i) = c_{Ri}^2 u^2(\delta \Delta R_i) + c_{Ii}^2 u^2(\delta \Delta I_i) \text{ with } c_{Ri} = (P_i / R_i), c_{Ii} = (P_i / I_i).$$

Table 1. Typical uncertainties budget for working standard 0.001 Ω at 50 A

Quantity	Estimate of quantity		Standard uncertainty		Coefficient of sensitivity	Uncertainty contribution	
R_E	0.000 100 008 6	Ω	0.000 000 000 6	Ω	10.0	0.006 0	$\mu\Omega$
R_{Edrift}	0.000 000 000	Ω	0.000 000 000	Ω	1.0	0.000 0	$\mu\Omega$
R_{Etemp}	0.000 000 000	Ω	0.000 000 000 01	Ω	1.0	0.000 0	$\mu\Omega$
R_{Xtemp}	0.000 000 000	Ω	0.000 000 001	Ω	1.0	0.000 8	$\mu\Omega$
r_c	1.000 000 000		0.000 006 250		0.001 Ω	0.006 3	$\mu\Omega$
R	10.001 333 785		0.000 003 568		0.000 1 Ω	0.000 4	$\mu\Omega$
R_x	0.001 000 219 2	Ω				0,008 7	$\mu\Omega$
					for k = 2	0,017 4	$\mu\Omega$

Table 2. Typical expanded uncertainties for air cooled shunts in lab (± 1 °C) and example of typical uncertainty budget for 0.005 Ω shunt at 100 A

Nominal value	Current	Expanded uncertainty (k = 2)
0.1 Ω	5 A	6.3 ppm
	10 A	6.3 ppm
	10 A	10.8 ppm
	20 A	10 ppm
0.01 Ω	40 A	19.8 ppm
	40 A	21.9 ppm
	70 A	15.4 ppm
	100 A	15.3 ppm

Quantity	Estimate of quantity		Standard uncertainty		Coefficient of sensitivity	Uncertainty contribution	
R_E	0.001 000 224	Ω	0.000 000 005	Ω	5.08	0.025	$\mu\Omega$
R_{Edrift}	0.000000000	Ω	0.000 000 000	Ω	1.0	0.000	$\mu\Omega$
R_{Etemp}	0.000000000	Ω	0.000 000 000	Ω	1.0	0.000	$\mu\Omega$
R_{Xtemp}	0.000000000	Ω	0.000 000 029	Ω	1.0	0.029	$\mu\Omega$
r_c	1.000000000		0.000 001 091		0.005 Ω	0.005	$\mu\Omega$
R	5.080671880		0.000 001 740		0.001 Ω	0.002	$\mu\Omega$
R_x	0.005 081 811	Ω				0.039	$\mu\Omega$
					for k = 2	0.078	$\mu\Omega$

Table 3. Typical expanded uncertainties for air cooled shunts in thermostatic box (± 0.15 °C) and example of typical uncertainty budget for 0.005 Ω shunt at 10 A and 30 °C

Nominal value	Current	Expanded uncertainty (k = 2)
0.1 Ω	1 A	2 ppm
0.01 Ω	5 A	8.4 ppm
	10 A	8.0 ppm
0.008 5 Ω	10 A	8.9 ppm
0.005 Ω	10 A	10.4 ppm

Quantity	Estimate of quantity		Standard uncertainty		Coefficient of sensitivity	Uncertainty contribution	
R_E	0.020 000 577	Ω	0.000 000 041	Ω	0.254	0.018	$\mu\Omega$
R_{Edrift}	0.000 000 000	Ω	0.000 000 000	Ω	1.0	0.000	$\mu\Omega$
R_{Etemp}	0.000 000 000	Ω	0.000 000 007	Ω	1.0	0.007	$\mu\Omega$
R_{Xtemp}	0.000 000 000	Ω	0.000 000 017	Ω	1.0	0.017	$\mu\Omega$
r_c	1.000 000 000		0.000 000 873		0.005 Ω	0.004	$\mu\Omega$
R	0.254 087 795		0.000 000 057		0.02 Ω	0.001	$\mu\Omega$
R_x	0.005 081 903	Ω				0.0263	$\mu\Omega$
					for k = 2	0.053	$\mu\Omega$

Typical values of TCR and PCR

TCR: 18°C – 30°C

Foil shunts: -2.8 +0.2 +8 ppm, unc. < 2.1ppm

Shunts with parallel rez.: +1.2 +1.7 ppm, unc. < 1.7 ppm

PCR: 50% - 100% of nominal current

Foil shunts: < ± 4 ppm, unc. < 3.1ppm

Shunts with parallel rez.: < ± 1.5 ppm, unc. < 1.5 ppm

Conclusions

Most significant influences for DCR (except of working standard calibration): temperature influence on air cooled shunts and/or standard deviation of measured voltages ratio.

Future work focus on trying to reduce the uncertainties - especially the contribution of measured ratio (standard deviation).

Acknowledgements

Authors special thank Mr. Martin Garcocz (BEV), Valter Tarasso (SP) and Matjaz Lindic (SIQ) for cooperation and shunts lending.

The research presented in this paper is part of the EURAMET joint research project on "Power and Energy" and has received funding from the European Community's Seventh Framework Programme, ERA-NET Plus, under Grant Agreement No. 217257.

References

- [1] V. Novakova Zachovalova, L. Indra, M. Sira, "Measurement System for DC Characterization of Low Resistance Standards," NCSL Int. Conf. Proc.2009, San Antonio, USA, July 26-30, 2009.
- [2] M. Garcocz, P. Scheibenreiter, W. Waldmann and G. Heine: "Expanding the measurement capability for AC-DC Current Transfer at BEV", CPEM 2004 Digest, pp. 461-462.
- [3] K. -E. Rydler, V. Tarasso, "Extending ac-dc current transfer measurement to 100 A, 100 kHz," CPEM 2008 Digest, pp. 28-29.
- [4] V. Novakova Zachovalova, "AC-DC current transfer difference in CMI," CPEM 2008 Digest, pp. 362-363.
- [5] B. Voljc, M. Lindic, R. Lapuh, "Direct Measurement of AC Current by Measuring the Voltage Drop on the Coaxial Current Shunt," IEEE Trans. on Instr. and Meas., vol. 58, no. 4, pp. 863-867, 2009.